



ELISA Thrombin-Heparin Cofactor II-Complex

Capture and detecting antibodies for the construction of an ELISA for the quantitative determination of Thrombin-HCII-complex

Test principle

Micro wells are coated with anti-thrombin-antibodies (capture antibody), to which Thrombin-HCII-complex from the sample binds. Unbound substances are removed by washing. A peroxidase conjugated detecting antibody is added and binds to the Thrombin-HCII-complex/capture antibody complex. After a washing step, added enzyme substrate reacts with the conjugate. The intensity of the resulting color is proportional to the Thrombin-HCII-complex content of the sample.

Reagents and stability (after opening)

Capture antibody 1 vial

Detecting antibody 1 vial

When stored at -20°C the reagents remain stable until the expiry date printed on the vials.

No. of determinations: 5 x 96

Article no.: THCII-EIA