



UNITEST Factor XIa Inhibitors

A chromogenic method for the determination of factor XIa inhibitors in plasma.

Test principle

FXIa-INH + FXIa (excess) →

[FXIa-INH • FXIa] + FXIa (residual)

Substrate $\xrightarrow{\text{Faktor XIa (residual)}}$ Peptide + pNA

Reagents and stability (after opening)

Factor XIa substrate	1 vial	6 months at 4°C
Factor XIa	1 vial	6 hours at 4°C
Human albumin	1 vial	8 hours at 4°C
Buffer concentrate	1 vial	see expiry date
Standard plasma	1 vial	8 hours at 4°C

Article no.: 0091