



## ELISA Factor XII

---

Capture and detecting antibodies for construction of an ELISA for the quantitative determination of factor XII antigen.

### Test principle

Micro wells are coated with factor XII antibodies (capture antibody), to which factor XII from the sample binds. Unbound substances are removed by washing. A peroxidase conjugated detecting antibody is added and binds to the factor XII/capture antibody complex. After a washing step, added enzyme substrate reacts with the conjugate. The intensity of the resulting color is proportional to the factor XII content of the sample.

### Reagents and stability (after opening)

Capture antibody	1 vial
Detecting antibody	1 vial

When stored at 2-8°C the reagents remain stable until the expiry date.

**No. of determinations:** 5 x 96

**Article no.:** FXII-EIA