



## ELISA Vitronectin

Complete ELISA kit for the quantitative determination of human vitronectin.

### Test principle

The samples are incubated in micro wells coated with an anti-vitronectin antibody. After binding of vitronectin to the capture antibody, unbound substances are removed by washing. A peroxidase conjugated detecting antibody is added and binds to vitronectin /capture antibody complex. After a washing step, added substrate reacts with the conjugate. The intensity of the resulting colour is proportional to the content of vitronectin in the sample.

### Reagents and stability (after opening)

Micro wells	12x8	8 weeks at 4°C
Diluent buffer	3 vials	4-8 weeks at 4°C
Calibrator	1 vial	4 hours at RT 8 weeks at -70°C
Conjugate	1 vial	4 hours at RT
Wash buffer (12.5 x)	1 vial	4-8 weeks at 4°C
TMB-substrate	1 vial	see expiry date at 2-8°C
Stop solution	1 vial	see expiry date at 2-8°C

**No. of determinations:** 96

**Article no.:** TC12120