



## VisuLize® Factor IX Antigen Kit

Complete ELISA Kit for the quantitative determination of factor IX Antigen. For in vitro diagnostic use for determination of factor IX in plasma and concentrates.

### Test principle

Micro wells are coated with factor IX antibodies (capture antibody), to which factor IX from the sample binds. Unbound substances are removed by washing. A peroxidase conjugated detecting antibody is added and binds to the factor IX/capture antibody complex. After a washing step, added enzyme substrate reacts with the conjugate. The intensity of the resulting colour is proportional to the content of factor IX in the sample.

### Reagents and stability (after opening)

Micro wells	6 x 16	expiry date
Calibrator	2 vials	4 hours at RT 30 days at -20°C
Control plasma A	2 vials	4 hours at RT 30 days at -20°C
Control plasma B	2 vials	4 hours at RT 30 days at -20°C
Wash buffer	1 vial	expiry date
Sample buffer	3 vials	expiry date
Conjugate	1 vial	expiry date
Substrate	1 vial	expiry date
Stop solution	1 vial	expiry date

**No. of determinations:**

**96**

**Article no.:**

**FIX-AG**