



## ELISA Factor IX

---

Capture and detecting antibodies for construction of an enzyme immuno assay for quantitative determination of factor IX.

### Test principle

Micro wells are coated with factor IX antibodies (capture antibody), to which factor IX from the sample binds. Unbound substances are removed by washing. A peroxidase conjugated detecting antibody is added and binds to the factor IX/capture antibody complex. After a washing step, added enzyme substrate reacts with the conjugate. The intensity of the resulting colour is proportional to the content of factor IX in the sample.

### Reagents and stability (after opening)

Capture antibody 1 vial  
Detecting antibody 1 vial  
When stored at -20°C the reagents are stable until stated expiry date.

**No. of determinations:** 5 x 96

**Article no.:** (human) FIX-EIA

**No. of determinations:** 4 x 96

**Artikel No.:** (canine) CFIX-EIA