



## ELISA Glu-plasminogen

Complete ELISA-kit for a quantitative determination of Glu-plasminogen in human plasma.

### Test principle

The diluted samples are incubated in micro wells coated with monoclonal anti-plasminogen-antibodies. After binding of glu-plasminogen to the capture antibody, unbound substances are removed by washing. A peroxidase conjugated detecting antibody is added and binds to the glu-plasminogen/capture antibody complex. After a washing step, added substrate reacts with the conjugate. The intensity of the resulting colour is proportional to the content of glu-plasminogen in the sample.

### Reagents and stability (after opening)

Coated micro wells	12x8	expiry date at 2-8°C
Wash buffer (10-fold)	1 vial	6 months at 2-8°C
Incubation buffer	1 vial	2 months at 2-8°C
Standard	1 vial	6 months at -20°C
Conjugate	1 vial	6 months at 2-8°C
TMB-substrate	1 vial	until exp.date at 2-8°C
Stop solution	1 vial	expiry date

**No. of determinations:** 96

**Article no.:** TC12040