



Technoclot® PT Owren manual

Coagulation assay for the quantitative determination of Prothrombin Time (PT) in human citrated plasma. The lyophilised reagent is sensitive to abnormal levels of coagulation factors II, VII and X and is used for the monitoring of oral anticoagulant therapy (i.e. warfarin). The reagent contains rabbit brain thromboplastin and adsorbed bovine plasma as a source for the coagulation factors V and fibrinogen and CaCl₂.

Technoclot® PT Owren replaces *Thrombotest und Normotest*.

Test principle

Prewarm the reagent at 37°C and add the sample. Start timing of the coagulation.

Reagents and stability (after reconstitution)

Technoclot PT Owren manual	1 hour at 37°C
	6 hours at RT
	24 hours at 2-8°C
	1 month at -20°C

Package size: 10 x 4 ml

Article no.: 5005032

Package size: 10 x 10 ml

Article no.: 5005037